

STANDARD OPERATING PROCEDURE

Calibration & Liquid Permittivity Extraction for SIW C-Band Resonators

DOCUMENT TITLE:	SIW C-Band Resonator SOP	DOC NUMBER:	SOP-SIW-001
TARGET HARDWARE:	Substrate Integrated Waveguide / C-Band (4-8 GHz) Fluidic Cell	REVISION / DATE:	Rev 1.0 / Sep 16, 2026
AUTHOR / SPECIALIST:	Kadeem Manning (Lead Field Specialist)	STATUS:	APPROVED / FIELD READY

1. Purpose & Scope

This Standard Operating Procedure (SOP) defines Vector Network Analyzer (VNA) 2-port calibration, microfluidic channel loading, and complex permittivity extraction (ϵ_r' , ϵ_r'') for Substrate Integrated Waveguide (SIW) and Substrate Integrated Suspended Line (SISL) C-band resonator sensors. SIW resonators provide high Quality Factors ($Q_0 > 250$) and electromagnetic shielding for high-precision dielectric characterization of liquids.

2. Required Test Equipment & Calibration Standards

- 2-Port Vector Network Analyzer (VNA):** Keysight FieldFox N9918A (300 kHz to 26.5 GHz).
- VNA Calibration Kit:** Precision N-type 50-ohm SOLT (Short-Open-Load-Thru) mechanical calibration kit.
- Precision Torque Wrench:** 3.5 mm SMA connector torque wrench (5.0 in-lbs $\pm$ 0.2 in-lbs).
- Reference Calibration Liquids:** Deionized water ($\epsilon_r' = 78.4$ at 25°C), Ethanol ($\epsilon_r' = 24.5$), and Isopropanol ($\epsilon_r' = 18.3$).
- Microfluidic Syringe Pump:** Syringe driver (0.1–10.0 mL/min flow rate) and PTFE tubing.

3. Step-by-Step Calibration & Permittivity Extraction Procedure

3.1 VNA 2-Port SOLT Calibration & Thermal Stabilization

- Power on VNA and allow 30 minutes for internal reference oscillator thermal stabilization.
- Set frequency sweep range: 4.0 GHz to 8.0 GHz (1601 points, 100 Hz IF bandwidth, 0 dBm power).
- Connect precision SOLT standards to Port 1 and Port 2 cables. Execute 2-port SOLT calibration. Verify S_{11} , $S_{22} < -45$ dB across band.

3.2 Unloaded Air Baseline Verification (Q_0 & f_0)

- Connect SIW resonator sensor board to VNA test cables using torque wrench (5.0 in-lbs).
- Measure S_{21} transmission response with empty microfluidic channel.
- Record unloaded resonant frequency $f_0 \approx 5.80$ GHz and peak $S_{21} \geq -3.5$ dB. Calculate 3-dB bandwidth Δf_0 and unloaded Quality Factor $Q_0 = f_0 / \Delta f_0$. Confirm $Q_0 \geq 250$.

3.3 Liquid Sample Injection & Dielectric Inversion

- Inject 15.0 mL of reference liquid into the SIW etched slot channel at 2.0 mL/min. Ensure zero air bubbles in sensing region.
- Record shifted resonant frequency f_r and loaded Quality Factor Q_L .
- Calculate real permittivity ϵ_r' and loss tangent $\tan \delta$ using perturbation equations:
$$\frac{f_0 - f_r}{f_0} = A (\epsilon_r' - 1) \quad \text{and} \quad \Delta \left(\frac{1}{Q} \right) = B \cdot \epsilon_r''$$

4. Quality Control & Temperature Correction

Substrate permittivity changes at $-0.0015/^{\circ}\text{C}$. Measure ambient temperature during sweep and apply quadratic thermal compensation to f_r . Flush microfluidic channel with methanol between samples and verify f_0 returns to within ± 2 MHz of air baseline.